

Diaphragm Seals

DS/PI-49A

ASHCROFT DIAPHRAGM SEALS ON CHLORINE APPLICATIONS

Halocarbon fill fluid must be used on all chlorine applications. If the process media includes chlorine in any form, Halocarbon is required. Examples are:

- Chlorine (Dry)
- Chlorine (Moist)
- Chlorine Vapors
- Chlorine Dioxide
- Chlorine Trifluoride
- Etc.

If a liquid filled gauge is attached to the seal, the fill fluid must again be Halocarbon.

Consult corrosion data guide for material recommendations. If chemicals are not listed, consult Customer Service at 203-378-8281.